

# Summary Report of IQC program for G6PD Quantitative Test - AMP Group - June 2015 -

## I. The statistic results of all laboratories in this month

G6PD	Control N (Lot No.:AC1203N)	Control D (Lot No.:AC1203D)
Labs	14	14
Received results number (n)	85	85
Median	12.6 (U/gHb)	3.1 (U/gHb)
Mean	12.7 (U/gHb)	3.1 (U/gHb)
SD	1.6	0.5
CV	12.6%	16.1%
Range of G6PD	9.4 ~ 15.6 (U/gHb)	2.2 ~ 4.5 (U/gHb)
Range of Hb	1.9 ~ 3.1 (g/dL)	1.5 ~ 5.1 (g/dL)

\*The statistic results are calculated from all labs reported in this month

## II. The distribution of G6PD reported for each lab in this survey

