

Summary Report of IQC program for G6PD Quantitative Test - AMP Group - September 2015 -

I. The statistic results of all laboratories in this month

G6PD	Control N (Lot No.:AC1203N)	Control D (Lot No.:AC1203D)
Labs	13	13
Received results number (n)	59	59
Median	12.7 (U/gHb)	3.1 (U/gHb)
Mean	12.9 (U/gHb)	3.1 (U/gHb)
SD	1.3	0.4
CV	10.1%	12.9%
Range of G6PD	9.9 ~ 15.5 (U/gHb)	2.0 ~ 4.0 (U/gHb)
Range of Hb	1.6 ~ 2.8 (g/dL)	2.4 ~ 3.5 (g/dL)

*The statistic results are calculated from all labs reported in this month

II. The distribution of G6PD reported for each lab in this survey

