

# Summary Report of IQC program for G6PD Quantitative Test - AMP Group - April 2016 -

## I. The statistic results of all laboratories in this month

G6PD	Control N (Lot No.:AC1203N)	Control D (Lot No.:AC1203D)
Labs	10	10
Received results number (n)	51	51
Median	12.0 (U/gHb)	2.8 (U/gHb)
Mean	11.9 (U/gHb)	2.8 (U/gHb)
SD	1.2	0.2
CV	10.1%	17.9%
Range of G6PD	9.2 ~ 14.3 (U/gHb)	1.7 ~ 3.9 (U/gHb)
Range of Hb	2.2 ~ 3.9 (g/dL)	1.3 ~ 3.5 (g/dL)

\*The statistic results are calculated from all labs reported in this month

## II. The distribution of G6PD reported for each lab in this survey

