

Summary Report of IQC program for G6PD Quantitative Test - AMP Group

- May 2016 -

I. The statistic results of all laboratories in this month

G6PD	Control N (Lot No.:AC1203N)	Control D (Lot No.:AC1203D)
Labs	14	14
Received results number (n)	66	66
Median	12.0 (U/gHb)	2.8 (U/gHb)
Mean	12.1 (U/gHb)	2.8 (U/gHb)
SD	1.1	0.3
CV	9.1%	10.7%
Range of G6PD	9.2 ~ 14.6 (U/gHb)	2.0 ~ 3.6 (U/gHb)
Range of Hb	1.6 ~ 2.7 (g/dL)	1.9 ~ 3.2 (g/dL)

*The statistic results are calculated from all labs reported in this month

II. The distribution of G6PD reported for each lab in this survey

