

Summary Report of IQC program for G6PD Quantitative Test - AMP Group - July 2016 -

I. The statistic results of all laboratories in this month

G6PD	Control N (Lot No.:AC1203N)	Control D (Lot No.:AC1203D)
Labs	10	10
Received results number (n)	46	46
Median	1.9 (U/gHb)	2.9 (U/gHb)
Mean	12.2 (U/gHb)	2.9 (U/gHb)
SD	1.3	0.4
CV	10.7%	13.8%
Range of G6PD	10.3 ~ 14.5 (U/gHb)	1.8 ~ 3.6 (U/gHb)
Range of Hb	2.0 ~ 2.8 (g/dL)	2.3 ~ 3.2 (g/dL)

*The statistic results are calculated from all labs reported in this month

II. The distribution of G6PD reported for each lab in this survey

