

# Summary Report of IQC program for G6PD Quantitative Test - AMP Group - October 2016 -

## I. The statistic results of all laboratories in this month

G6PD	Control N (Lot No.:AC1203N)	Control D (Lot No.:AC1203D)
Labs	8	8
Received results number (n)	46	46
Median	12.0 (U/gHb)	2.8 (U/gHb)
Mean	11.9 (U/gHb)	2.8 (U/gHb)
SD	0.9	0.3
CV	7.5%	10.7%
Range of G6PD	10.2 ~ 13.6 (U/gHb)	2.1 ~ 3.4 (U/gHb)
Range of Hb	2.3 ~ 2.6 (g/dL)	2.5 ~ 3.1 (g/dL)

\*The statistic results are calculated from all labs reported in this month

\*\* G6PD Method = AMP reagent kit, 37°C

## II. The distribution of G6PD reported for each lab in this survey

