

Summary Report of IQC program for G6PD Quantitative Test - AMP Group - November 2016 -

I. The statistic results of all laboratories in this month

G6PD	Control N (Lot No.:AC1203N)	Control D (Lot No.:AC1203D)
Labs	8	8
Received results number (n)	33	33
Median	11.8 (U/gHb)	2.8 (U/gHb)
Mean	11.7 (U/gHb)	2.8 (U/gHb)
SD	0.8	0.3
CV	6.8%	10.7%
Range of G6PD	10.0 ~ 13.2 (U/gHb)	2.1 ~ 3.4 (U/gHb)
Range of Hb	2.1 ~ 2.9 (g/dL)	2.6 ~ 3.2 (g/dL)

*The statistic results are calculated from all labs reported in this month

** G6PD Method = AMP reagent kit, 37°C

II. The distribution of G6PD reported for each lab in this survey

