

Summary Report of IQC program for G6PD Quantitative Test - Medicon Group - October 2018 -

I. The statistic results of all laboratories in this month

G6PD	Control N (Lot No.:AE0909N)	Control D (Lot No.:AE0909D)
Labs	8	8
Received results number (n)	41	41
Median	15.0 (U/gHb)	5.3 (U/gHb)
Mean	14.9 (U/gHb)	5.3 (U/gHb)
SD	1.0	0.3
CV	6.7%	5.7%
Range of G6PD	12.3 ~ 18.0 (U/gHb)	4.7 ~ 6.2 (U/gHb)
Range of Hb	1.7 ~ 2.1 (g/dL)	2.0 ~ 2.4 (g/dL)

*The statistic results are calculated from all labs reported in this month

** G6PD Method = Medicon reagent kit, 37°C

II. The distribution of G6PD reported for each lab in this survey

