
## Summary Report of IQC program for G6PD Quantitative Test - Medicon Group - October 2023 -


## I. The statistic results of all laboratories in this month


G6PD	Control N (Lot No.:BJ0922N)	Control D (Lot No.: BJ0922D)
Labs	6	6
Received results number (n)	31	31
Median	14.9 (U/g Hb)	4.9 (U/g Hb)
Mean	14.8 (U/g Hb)	4.9 (U/g Hb)
SD	0.5	0.5
CV	3.4%	10.2%
Range of G6PD	$13.7 \sim 15.6  (\text{U/g Hb})$	$4.0 \sim 5.6 \; (U/g \; Hb)$
Range of Hb	$2.1 \sim 2.6  (g/dL)$	$1.9 \sim 2.3 \; (g/dL)$

<sup>\*</sup>The statistic results are calculated from all labs reported in this month

## II. The distribution of G6PD reported for each lab in this survey









<sup>\*\*</sup> G6PD Method = Medicon reagent kit, 37°C