
## Summary Report of IQC program for G6PD Quantitative Test - Medicon Group - November 2023 -


## I. The statistic results of all laboratories in this month


G6PD	Control N (Lot No.:BJ0922N)	Control D (Lot No.: BJ0922D)
Labs	6	6
Received results number (n)	33	33
Median	15.0 (U/g Hb)	4.9 (U/g Hb)
Mean	14.9 (U/g Hb)	4.9 (U/g Hb)
SD	0.4	0.4
CV	2.7%	8.2%
Range of G6PD	$14.1 \sim 15.5 \; (\text{U/g Hb})$	$4.2 \sim 5.5 \; (\text{U/g Hb})$
Range of Hb	$2.1 \sim 2.5 \text{ (g/dL)}$	$1.9 \sim 2.2  (g/dL)$

<sup>\*</sup>The statistic results are calculated from all labs reported in this month

## II. The distribution of G6PD reported for each lab in this survey









<sup>\*\*</sup> G6PD Method = Medicon reagent kit, 37°C