

Summary Report of IQC program for G6PD Quantitative Test - Medicon Group - September 2024 -

I. The statistic results of all laboratories in this month

G6PD	Control N (Lot No.:BJ0922N)	Control D (Lot No.: BJ0922D)
Labs	9	9
Received results number (n)	64	64
Median	14.8 (U/g Hb)	5.2 (U/g Hb)
Mean	14.7 (U/g Hb)	5.1 (U/g Hb)
SD	0.6	0.3
CV	4.1%	5.9%
Range of G6PD	13.4 ~ 15.6 (U/g Hb)	4.4 ~ 5.7 (U/g Hb)
Range of Hb	1.9 ~ 2.7 (g/dL)	1.9 ~ 2.7 (g/dL)

*The statistic results are calculated from all labs reported in this month

** G6PD Method = Medicon reagent kit, 37°C

II. The distribution of G6PD reported for each lab in this survey

